




E&I UltraX-10

Broadband Power Amplifier



Do not attempt to operate this unit prior to reading this manual.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

August 2023

Rev. A

Warranty

Electronics & Innovation Ltd., (hereafter E&I) warrants for the period of three year from the date of original delivery, each unit to be free of defects in materials and workmanship. For the period of 36 months E&I will, at its option, repair or replace defective parts so as to render the unit fully operational such that it performs according to the original specifications; free of charge to the original purchaser.

Should warranty service be required, the unit must be returned to E&I, freight cost to be borne by the owner. If, in our opinion, the unit has been damaged by use outside the limits prescribed in this manual or by accident, then the warranty shall not be honored. In such a case E&I will provide an estimate for repair, assuming repair is possible and provide a quote at standard service rates.

Contents

Chapter 1 General Information	3
Chapter 2 Operation	4
Chapter 3 Technical Information	5
Chapter 4 Maintenance	7
Chapter 5 Safety	10

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

Chapter 1 Introduction

The UltraX-10 is a broadband solid state amplifier covering the frequency spectrum from 100kHz to 5MHz. It is rated at 10 watts of RF power with low harmonic and intermodulation distortion. Over 20 watts of saturated power can be produced with increased distortion products. A highly linear Class A design, the UltraX-10 will amplify inputs of AM, FM, SSB, pulse and any complex modulation signals. The amplifier has 40 dB gain it is unconditionally stable and will not oscillate even with combinations of mismatched source and load impedance. It is protected against failure due to output load mismatch. An integral power supply permits operation from 115/230 single phase AC power.

1.2 INSTRUMENTATION IDENTIFICATION

Each amplifier is identified by a serial number tag on the back panel of the unit. Both the model number and the serial number should be quoted to identify specific unit.

1.3 SPECIFICATIONS

Physical and electrical specifications are listed in Table 1-1 below

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

Table 1-1. SPECIFICATIONS

100 kHz to 5 MHz
40 dB min, ±1.5 variation
Nominal 10 watts
> -20dBc at 10 watts output
50 ohms, VSWR, 1.5:1 Maximum
50 ohms, VSWR, 2:1 Maximum
Continuous operation into any load or source impedance
Unit will withstand a + 13dBm input signal (1.0 Volts RMS max) into 50 ohms. Unit will withstand +3dBm input for all load conditions
100 – 240 VAC 47-63 Hz.
3.25 x 8.25 x 11.63 inches 82.55 x 209.55 x 295.5 mm
5.5lbs (2.27kg)
BNC
0 – 40 C

Chapter 2 Operation

2.1 INTRODUCTION

The UltraX-10 RF amplifier is used to amplify the RF level of signal sources in the 100kHz to 2MHz range. No tuning or any other form of adjustment is required.

The UltraX-10 produces power output at its output connector, regardless of load impedance. Any power reflected due to output load mismatch is absorbed in the amplifier. Therefore, although the output impedance is 50 ohms (maximum VSWR: 2:1), the amplifier will work into any load impedance.

The maximum input applied should be less than +13 dBm (into 50ohm load), and +3dBm in any other load. Continuous operation above this level can cause degradation and failure.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

2.2 RACK INSTALLATION

This unit is 2U high, 12" width. With it will fit into a standard rack.

2.2.1 Mains Voltage

The unit accommodates AC line voltages from 100 TO 240 VAC 47 – 63 Hz

2.3 OPERATION

A line cord is supplied to form a connection between the mains supply and the rear of the unit. Plug this into AC input at the rear of the unit and the AC mains outlet.

2.3.1 Proceed as follows:

- (i) Ensure that there is at least 3" or 7.5 cm clearance at the rear of the unit for air flow.
- (ii) Switch on power and observe that the green led illuminates. (A brief flash of the fault led is normal at turn on).
- (iii) Ensure RF input voltage is not excessive
 - a. The 1 V rms indicated maximum input voltage is 5 times the level of the input signal required to achieve maximum output. Input voltages in excess of 2 volts peak may permanently damage the instrument.
- (iv) Connect the output via a 50 ohm coaxial lead and BNC plug to the load.
- (v) Connect the input signal via a 50 ohm coaxial lead and BNC plug to the input connector.


Chapter 3 Technical Description

3.1 GENERAL DESCRIPTION

The UltraX-10 is designed to amplify signals by 40dB in the frequency band of 100kHz to 5MHz. The signal from the front panel BNC connector is fed via a length of 50 ohm coaxial cable into the input of the two way splitter. The signal is split into two and fed to the two power amplifiers that are in parallel with each other. Each power amplifier is a two stage unit, each stage comprising of a push-pull LDMOS Fet. The output of each stage is then combined and fed to the output of the unit.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

The power supply unit provides a 35.3 VDC 5 ampere source. It is a switch mode power supply unit. The output feeds the PA The power supply also has a 5 VDC output which feeds the fault detection board.



If the PA heat sink temperature exceeds 70C the unit will fault to prevent damage. This will be indicated by a Red LED on the front panel. The unit will not need to be power cycled to reset the fault condition.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

August 2023

Rev. A

Chapter 4 Maintenance

4.1 INTRODUCTION

The UltraX-10 RF amplifier requires no periodic maintenance. The instrument is unconditionally stable and is fail-safe under all load conditions. Damage can only be externally caused by the incorrect selection of the AC supply voltage or by an input signal in excess of the specified 1-volt rms equivalent to a power level of +13dBm.

This chapter therefore, deals only with certain fundamental procedures for fault location.

Performance limits quoted are for guidance only and should not be taken for guaranteed performance specifications unless they are also quoted in the Specification Section 1.2.

4.2 PERFORMANCE CHECKS

To determine the amplifier's performance, carry out the following procedure.


4.2.1 Initial Check

The following check can be made after repair and adjustments or whenever the condition of the unit is in question.

- a. Connect AC power supply. Switch on power and observe that the green led illuminates. A brief flash of the fault led is normal at turn on.
- b. Connect a sweep generator (HP 8601 or similar) capable of sweeping the frequency range 50 KHz to 2 MHz, to the input connector.
- c. Adjust the output level of the sweep generator so that a 50 ohm video detector connected at the output of the unit will not be damaged by excessive power output. (Reference section 4.4.1 for set up.)
- d. Observe the gain versus frequency ripple on an oscilloscope calibrated in decibels. The gain variation must be not more than +/- 1.5 dB over the frequency range.
- e. Connect a calorimetric power meter (HP434 or equivalent) through a 10dB 200-watt attenuator to the output connector. Adjust the input CW signal to any frequency between 100 kHz and 2 MHz for 10 watts output.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

- f. Observe the harmonic distortion of the output on a spectrum analyzer. The harmonic components contributed by the amplifier should be better than 20 dB down from the fundamental.
 - (a) If the above items are found to be outside of the specification, check the spectral content of the input signal. If this is a pure signal then the unit needs to be returned to the factory for service.
- 4.4.2 Measurement of Harmonics





- 1.) Equipment Required:
 - a) Sweep/Signal Generator

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

- b) Calorimetric Power Meter HP434A
- c) Spectrum Analyzer.
- d) Attenuator (20dB)
- e) Attenuator (30dB)
- f) Coupler (30 dB)

2.) Connect the Equipment as shown in Figure 4-2, then proceed as follows:

- a) Adjust the signal generator to a CW center frequency of 1 MHz, for an indicated output of 10 watts on the power meter.
- b) Using the spectrum analyzer, check that the level of the carrier harmonics is less than -20 dB with respect to the carrier while manually scanning the frequency band of 100kHz to 2MHz. An indicated power output of 10W should be maintained during this operation.

4.5 PACKAGING FOR RESHIPMENT

In the event of the equipment being returned for servicing it should be packed in the original shipping carton and packing material. If this is not available, wrap the instrument in heavy paper or plastic and place in a rigid outer box of wood, fiberboard or very strong corrugated cardboard. Use ample soft packing to prevent movement. Provide additional support for projecting parts to relieve these of unnecessary shock. Close the carton securely and seal with durable tape. Mark the shipping container FRAGILE to ensure careful handling.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.

Chapter 5 Safety:

Do not attempt to operate this unit with the cover removed. High AC and DC voltages are present. The cover prevents against electrical shock due to AC line voltages, high DC and RF fields. Further the cover provides part of the cooling system design. Components, specifically on the RF driver board are prone to overheat and eventual failure if the unit is operated without the cover in place.

Ensure that the load is connected to the output prior to connecting the RF input to the unit. This will prevent high voltages being present and exposed at the output connector.

Only use the AC cord provided or equivalent.

Ensure that the mains outlet is properly grounded.

The material contained in this document is the property of Electronics & Innovation, Ltd., it is subject to change without notice.